PORTABLE LED HEADLIGHT SYSTEM




DOVER™

EXAM: MEDIUM INTENSITY

PORTABLE LED HEADLIGHT WITH FIXED SPOT, **POWERED BY A 3-CELL BATTERY**

- Illumination Range: ~30,000 LUX at 16" (40 cm) to ~105,000 LUX at 10" (25 cm)
- Fixed Spot Diameter: 3" (7.6 cm) at a working distance of 16" (40 cm)
- Color Temperature: 7,000 K
- Intensity Settings: 5 Levels
- 3-Cell Battery
- Battery Duration at Full Intensity: 10 hours
- Battery Expected Life Span: > 300 Charges
- LED Expected Life Span: > 10,000 hours
- Optic Weight: 0.45 oz (12.7 g)
- Battery Weight: 6.5 oz (184 g)





DOVER PORTABLE LED HEADLIGHT SYSTEM INCLUDES:

- Dover High-Intensity Headlight
- Choice of: Fiji or Key West Headband, or Optic Clip
- Choice of One or Two 3-Cell Battery
- · Charging Cradle, Power Supply
- · Holster, Linkage Kit

DOVER PORTABLE LED HEADLIGHT SYSTEM PRODUCT CODES:

BFW 9030 Dover Headlight System with Fiji Headband with One 3-Cell Battery Dover Headlight System with Fiji Headband and Two 3-Cell Batteries BFW 9030B BFW 9032 Dover Headlight System with Optic Clip with One 3-Cell Battery BFW 9032B Dover Headlight System with Optic Clip and Two 3-Cell Batteries Dover Headlight System with Key West Headband with One 3-Cell Battery BFW 9034

Dover Headlight System with Key West and Two 3-Cell Batteries BFW 9034B

BFW 1552C Dover Optic Clip Adapter for Face Shields and Loupes

BFW 5440 Dover Protective Carrying Case, molded interior provides space for all components

BFW 9022ACC **Dover Charger** BFW 9035 **Dover 3-Cell Battery** BFW 9036 **Dover Holster**

EXCELLENT FOR USE IN THE FOLLOWING DEPARTMENTS

HOSPITAL

ENT GENERAL SURGERY CARDIOSURGERY **NEUROSURGERY ORTHOPEDIC SURGERY** PEDIATRIC SURGERY **SPINE SURGERY** THORACIC SURGERY TRANSPLANT SURGERY VASCULAR SURGERY

SURGERY CENTER

OFFICE

ORTHOPEDIC SURGERY PLASTIC SURGERY

DENTAL **EXAM**

ORAL/MAXILLOFACIAL



800-717-4673 | +1-502-899-1808

BFWinc.com

2843 SAMCO ROAD, SUITE A

RAPID CITY, SD 57702 USA

ENGINEERED AND ASSEMBLED IN THE U.S.A. © 2024 BFW. Inc.

ALL TECHNICAL SPECIFICATIONS ARE SUBJECT TO CHANGE WITHOUT NOTICE

DOVER TESTED AND CERTIFIED TO: IEC 60601-1 Edition 3, IEC 60601-1-2 Edition 4, ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10), CAN/CSA C22.2 NO 60601-1 Edition 2 (2008), CE